



MAKERS OF RESPIRATORY VALVES SINCE 1938

HANS RUDOLPH, inc.

www.rudolphkc.com

TEL: (913) 422-7788 U.S.A. & CANADA (800) 456-6695
 FAX: (913) 422-3337 E-Mail: hri@rudolphkc.com
 8325 COLE PARKWAY, SHAWNEE, KANSAS 66227 U.S.A.

Data Sheet...

RUDOLPH VOLUME CALIBRATION SYRINGES

Series 5520, 5510, 5550, 5540, 5530, 5570, 4900

Our **Calibration Syringes**, having a known calibrated volume as a standard, provide an easy and reliable method of calibrating and measuring the accuracy of respiratory volume measuring equipment in the field on a regular basis to satisfy your **quality assurance program**.

The 3 Liter size Syringe has been recommended as a minimum standard by the "Report of 1977 Snow Bird Workshop on Standardization of Spirometry", as the minimum volume to test spirometers. Our 7 liter size is the largest size Syringe, known to us, to be commercially available. This larger size is very useful in checking your Spirometer through a full 7 liters in a single uninterrupted complete volume stroke.

SPECIFICATIONS

Model Series	5520	5510	5550	5540	5530	5570	4900
Total Volume	10ml	100ml	1/2L	1L	3L	3L	7L
Smallest Graduation (ml)	2	20	50	50	100	3000	100
Volume Stroke (inch)	3	5	6 1/4	12.38	14 3/4	14 3/4	15
(cm)	7.62	12.70	15.78	31.45	37.47	37.47	38.10
Cylinder Bore (ID)(inch)	1/2	1 1/4	2 1/2	2 1/2	4	4	6
(cm)	1.27	3.17	6.35	6.35	10.16	10.16	15.24
Weight (lbs.)	.17	.39	1.7	2	5	5.3	10
(grams)	75	175	771	907	2268	2411	4536

ADJUSTABLE POSITIVE STOP FOR SETTING VOLUME

These stops allow accurate repetitive duplication of your selected volume.

1/2, 1, 3 & 7 Liter Syringes -

- The volume stop is easily located on the graduated piston rod for the desired volume.
- A socket-hex wrench is supplied for tightening the volume-stop
- The volume stop can be set throughout the entire 1/2, 1, 3 & 7 Liter range.
- The 5570 Series 3 Liter Syringe is factory set to 3L and is not adjustable.

100cc and 10cc Miniature Syringes -

- The method of setting the volume stop is unlike the 1/2, 1, 3 & 7 Liter Syringes.
- Each volume graduation on the piston rod has a groove which engages with a spring loaded detent, in the volume stop, for locating the desired volume. A screw, in the volume stop, is then tightened with a socket-hex wrench to securely hold the stop.
- These volume stops can only be set at the following graduations:
 5510 Series - 20ml, 40ml, 60ml, 80ml, & 100ml
 5520 Series - 2ml, 4ml, 6ml, 8ml, & 10ml

OUTLET PORT ADAPTERS

- They thread into the end plate, making it convenient to use the various standards or custom OD / ID size connections. Some 5570 Series 3L Syringe and some 5530 Series Syringes have integrally molded outlet adapters and are not removable.
- The 1/2, 1, 3 & 7 liter sizes have the same thread and are therefore interchangeable.
- **The Miniature 100ml and 10ml Model end plates have different size threads for the outlet port adapters, therefore these outlet port adapters can NOT be interchanged.**



QUALITY ASSURANCE

- **Traceable to the National Institute of Standards and Technology (NIST).**
- The accuracy of each syringe is $\pm 1/2$ of 1% full scale.
- **Quality Assurance guidelines recommend that inspection and recalibration procedures of the syringe accuracy be performed three to four times a year, once a year at a minimum.**

CONSTRUCTION

- 1/2, 1, 3 & 7 Liter cylinders are aluminum with hard anodize.
- 100 & 10 ml cylinders are acrylic plastic.
- Piston Seals are PTFE or nitrile rubber.
- Piston rods are aluminum (black anodize), stainless steel or polyamide thermoplastic.
- End plates, port adapters are POM, PET or PBT plastic.

MAINTENANCE / SERVICE (RE-CALIBRATION)

- Lubrication of seal should only be required once a year for the typical lab.
- For any re-conditioning, recalibration, replacing piston seals or replacing end cap o-rings, we encourage your returning syringe to us.
- The cost is nominal. We will Quote you a firm price after an evaluation of your Syringe. Contact Hans Rudolph, inc. for a Return Authorization Number (RAN).
- A **Rental Program** is available if your system must stay up and running. Contact Hans Rudolph, inc office for information.
- **CAUTION:** United States law restricts this device to sale by or on the order of a physician.
- **WARNING:** Refer to Instructions for Use, Document No. 691203, shipped with product.

ORDERING INFORMATION

P/N-Model No.

5520 Series 10ml Syringe - 1/2"(1.3cm) Diameter Cylinder. Has 2ml graduations with positive adjustable locking stop. Full volume stroke is 3"(7.6cm)
112380-5520A 10ml Syringe with 200271 Outlet 15mmOD/10.5mmID tapers
112381-5520B 10ml Syringe with 200272 Outlet 22mmOD/15mmID tapers
Replacement Outlet Port Adapters for 10 ml Syringes
200271 Outlet Adapter 15mm OD/10.5mmID tapers
200272 Outlet Adapter 22mmOD/15mmID tapers

5510 Series 100ml Syringe - 1 1/4"(3.2cm) diameter cylinder. Has 20ml graduations with positive adjustable locking stop. Full volume stroke is 5" (12.7cm).
112378-5510A 100ml Syringe with 200269 Outlet, 15mmOD/10.5mmID tapers
112379-5510B 100ml Syringe with 200270 Outlet, 22mmOD/15mmID tapers
Replacement Outlet Port Adapters For 100ml Syringes
200269 Outlet Adapter 15mmOD/10.5mmID tapers
200270 Outlet Adapter 22mmOD/15mmID tapers

5550 Series 1/2L Syringe - 2 1/2"(6.35cm) diameter cylinder. Has 50 ml graduations with adjustable locking stop. Full volume stroke is 6 1/4"(15.78cm)
112742-5550 1/2 Liter Syringe with 200264 Outlet, 1 3/8"(35mm)OD/
1 1/8"(28.6mm)ID
112743-5550A 1/2 Liter Syringe with 200265 Outlet, 1 3/4"(44.5mm)OD/
1 1/8"(28.6mm)OD (**Long OD tapered outlet**)
112744-5550B 1/2" Liter Syringe with 200266 Outlet, 22mmOD/15mmID tapers
112745-5550C 1/2 Liter Syringe with 200267 Outlet, Fits over 2"(50.8mm)OD with
0-Ring slip fit
112746-5550D 1/2 Liter Syringe with 200268 Outlet, 1 1/8"(28.6mm)OD/22mmID
taper

5540 Series 1L Syringe - 2 1/2"(6.35cm) diameter cylinder. Has 50ml graduations with adjustable locking stop. Full volume is 12.38" (31.45cm).
112706-5540 1 Liter Syringe with 200264 Outlet, 1 3/8"(35mm)OD/
1 1/8"(28.6mm)ID
112707-5540A 1 Liter Syringe with 200265 Outlet 1 3/4"(44.5mm)OD/
1 1/8"(28.6mm)OD (**Long OD tapered outlet**)
112708-5540B 1 Liter Syringe with 200266 Outlet, 22mmOD/15mmID tapers
112709-5540C 1 Liter Syringe with 200267 Outlet, Fits over 2"(50.8mm)OD with
0-ring slip fit
112710-5540D 1 Liter Syringe with 200268 Outlet, 1 1/8"(28.6mm)OD/22mmID
taper

5530 Series 3L Syringe - 4"(10.2cm) diameter cylinder. Has 100 ml graduations with adjustable locking stop. Full volume stroke is 14 3/4"(37.5cm)
112028-5530 3 Liter Syringe with Integral Outlet 1 3/8" (35mm) OD/
1 1/8"(28.6mm)ID
112231-5530A 3 Liter Syringe with 200265 Outlet 1 3/4"(44.5mm)OD/1 1/8"
(28.6mm)OD (**Long OD tapered outlet**)
112232-5530B 3 Liter Syringe with 200266 Outlet 22mmOD/15mmID tapers
112233-5530C 3 Liter Syringe with 200267 Outlet Fits over 2"(50.8mm)OD with
0-ring slip fit
112567-5530D 3 Liter Syringe with Integral Outlet 1 1/8"(28.6mm)OD/22mmID
taper

5570 Series 3L Syringe - 4"(10.2cm) diameter cylinder. Factory set for 3.000 Liter volume stroke displacement - Not adjustable.
113391-5570 3 Liter Syringe with Integral Outlet 1 3/8" (35mm) OD/
1 1/8"(28.6mm)ID
113638-5570A 3 Liter Syringe with 200265 Outlet 1 3/4"(44.5mm)OD/1 1/8"
(28.6mm)OD (**Long OD tapered outlet**)
113611-5570B 3 Liter Syringe with 200266 Outlet 22mmOD/15mmID tapers
113392-5570D 3 Liter Syringe with Integral Outlet 1 1/8"(28.6mm)OD/22mmID
taper

4900 Series 7L Syringe - 6"(15.2cm) diameter cylinder. Has 100ml graduations with adjustable locking stop. Full volume stroke is 15"(38.1cm).
112092-4900 7 Liter Syringe with 200264 Outlet 1 3/8"(35mm)OD/1 1/8"
(28.6mm)ID
112234-4900A 7 Liter Syringe with 200265 Outlet 1 3/4"(44.5mm)OD/1 1/8"
(28.6mm)OD (**Long OD tapered outlet**)
112235-4900B 7 Liter Syringe with 200266 Outlet 22mmOD/15mmID tapers
112236-4900C 7 Liter Syringe with 200267 Outlet Fits over 2"(50.8mm)OD with
0-ring slip fit
112568-4900D 7 Liter Syringe with 200268 Outlet 1 1/8"(28.6mm)OD/22mmID
taper

Replacement Outlet Port Adapters for 1/2, 1, 3, & 7 Liter Syringes

200264 Outlet Adapter 1 3/8"(35mm)OD/ 1 1/8 "(28.6mm)ID
200265 Outlet Adapter 1 3/4"(44.5mm)OD/1 1/8"(28.6mm)OD taper (**Long OD tapered outlet**)
200266 Outlet Adapter 22mmOD/15mmID tapers
200267 Outlet Adapter 2"(50.8mm)ID 0-ring slip fits over 2"(50.8mm)OD
200268 Outlet Adapter 1 1/8" (28.6mm)OD/22mmID taper

Replacement Parts for Syringes
171171 'T' Allen Wrench. For Series 5520 and 5510
171125 'T' Allen Wrench. For Series 5550, 5540, and 4900
171129 'T' Allen Wrench. For Series 5530
171115 Setting Collar, stainless steel. For Series 5550, 5540, and 4900
200232 Setting Collar, ball plunger, stainless steel. For Series 5520 & 5510
171126 Knob. For Series 5550, 5540, and 4900
171168 Knob. For Series 5520 and 5510

Recalibration of your Syringe, Contact Hans Rudolph, inc. for Return Authorization Number (RAN):

955034 Recalibration of 4900 Series 7 Liter Calibration Syringe
955052 Recalibration of 5570 Series 3 Liter Calibration Syringe
955035 Recalibration of 5530 Series 3 Liter Calibration Syringe
955036 Recalibration of 5540 Series 1 Liter Calibration Syringe
955037 Recalibration of 5550 Series .5 Liter Calibration Syringe
955038 Recalibration of 5510 Series 100 ml Calibration Syringe
955039 Recalibration of 5520 Series 10 ml Calibration Syringe
955051 **Dead Space Measurement of New Syringe, (All Series)**
955050 **Recalibration and Dead Space Measurement of returned syringe.** Order at time of obtaining a Return Authorization Number (RAN). (All series). Certificate of calibration provided.

IN-COMING CALIBRATION REPORT - on returned calibration syringe. This report must be requested at time of obtaining Return Authorization Number (RAN). 5-7 Data points will be checked with pass or fail indications.

955312 In-Coming Calibration on 4900 series 7 liter calibration Syringe
955315 In-Coming Calibration on 5570 series 3 liter calibration Syringe
955311 In-Coming Calibration on 5530 Series 3 liter Calibration Syringe
955310 In-Coming Calibration on 5540 Series 1 liter Calibration Syringe
955309 In-Coming Calibration on 5550 Series .5 liter Calibration Syringe
955308 In-Coming Calibration on 5510 Series 100 ml Calibration Syringe
955307 In-Coming Calibration on 5520 Series 10 ml Calibration Syringe

CERTIFICATE OF CALIBRATION - Provides National Institute of Technology (NIST) traceability to Hans Rudolph, inc. syringe test standards and actual volume calibration data (5-7 points) on your syringe with stated accuracy.

955306 Certificate of Calibration on 4900 series 7 liter calibration Syringe
955316 Certificate of Calibration on 5570 series 3 liter calibration Syringe
955305 Certificate of Calibration on 5530 Series 3 liter Calibration Syringe
955304 Certificate of Calibration on 5540 Series 1 liter Calibration Syringe
955303 Certificate of Calibration on 5550 Series .5 liter Calibration Syringe
955302 Certificate of Calibration on 5510 Series 100 ml Calibration Syringe
955301 Certificate of Calibration on 5520 Series 10 ml Calibration Syringe

NIST TRACEABILITY of your Rudolph Volume Calibration Syringe with 5 to 10 times greater accuracy than standard certificate of calibration. Hans Rudolph, inc. will certify the volume displacement of any specified reading(s) per your request. The price includes standard certificate of recalibration/calibration. The price is per graduation mark (contact HRI for special pricing on 5 or more points per syringe).
955046 Certificate of Traceability to NIST (Syringe Volume Setting).

CLEANING

This product requires **special** cleaning and care, which is described in our **Instructions for Use 691203**.

APPLICABLE STANDARDS FOR MEDICAL TAPERED CONNECTIONS: 7.5mm, 10.5mm, 15mm, 22mm & 30mm

The 15mm, 22mm and 30mm tapered male and female diameter fitting connections are manufactured in accordance with the applicable standards as follows:

- ISO (The International Organization for Standardization) 5356-1 conical connectors, cones and sockets.

All 7.5mm and 10.5mm male and female diameters are also tapered. Hans Rudolph, inc. has designed and manufactured these two sizes with parameters similar to the 15mm and 22mm standards. Diameters other than the above five sizes are straight unless otherwise noted.

GENERAL INFORMATION

• **Packaged:** Individually polyethylene bagged and shipped clean, **non-sterile**.

REFERENCE

How to use a Volume Calibration Syringe for calibration checks of Volume Spirometers. Refer to Chapter Eight, Calibration, of the book titled *Office Spirometry, A Practical Guide to the Selection and Use of Spirometers*, by Paul L. Enright and Robert E. Hyatt, published by Lea & Febiger, Philadelphia PA USA 1987.